Verification of Compliance

No. EC.1282.0H140325.PSI0999 Test Report no. XMT0201401012P/LVD, XMT0201401012P/EMC

Certificate's Holder: Preekem Scientific Instruments Co., Ltd.

No. 2 Building, NO. 4299 Jindu Road, Xinzhuang

Industrial Area, Shanghai, China

Product: Microwave Digestion System

Model(s): WX-8000

Directives: 2006/95/EC Low Voltage

2004/108/EC Electromagnetic Compatibility

Standards: EN 61010-1:2010, EN 61000-6-1:2007,

EN 61000-6-3:2007, EN 61000-3-2:2006+A2:2009.

EN 61000-3-3:2013, BS EN 61326-1:2013,

BS EN 61326-2-1:2013

Remark: This Verification of Compliance has been issued on a voluntary basis. ECM confirms that a Technical Construction File (TCF) is existent for the above listed product(s). The TCF satisfactorily covers the essential requirements of the above listed Directive(s).

Other relevant Directives have to be observed in case they are applicable.

This Document is only valid for the equipment and configuration described and in conjunction with the TCF detailed above. Whereas the Manufacturer is responsible of the certification of the product(s) and not exempted to perform all the necessary activities before placing the product(s) on the market.

The Manufacturer is also responsible of the internal production control to ensure the product(s) are in compliance with the essential requirements of the above mentioned Directive(s).

This certificate can be checked for validity at www.entecerma.org

Date of issue MARCH 2014

Certification Chief Manager





Expiry date MARCH 2022

Certification Deputy Manager

